



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 & ANSI/NCSL Z540-1-1994

Braswell Scale & Equipment Co., Inc.

1180 Sweeten Creek Rd. Asheville, NC 28813
Daryl Farlow Phone: 828-274-3771

CALIBRATION

Valid to: January 27, 2013

Certificate Number: AC-1205

I. Mechanical

Table with 5 columns: PARAMETER / EQUIPMENT, RANGE, CALIBRATION AND MEASUREMENT CAPABILITY [EXPRESSED AS UNCERTAINTY(±)], REFERENCE STANDARD OR EQUIPMENT, METHOD(S). Rows include Class I Balances, Class II Balances, Class III Scales, and Class III L Scales.

Notes:

- 1. Calibration and Measurement Capabilities (Expanded Uncertainties) are based on approximately a 95% confidence interval, using a coverage of k=2.
2. This laboratory provides in-laboratory and field (on-site) calibrations. Since field (on-site) conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected in the field (on-site) than what is reported on the accredited scope.
3. This scope is part of and must be included with the Certificate of Accreditation No. 1205

Handwritten signature of Keith Greenaway

Vice President

